

Occupational Safety Research Institute, v.v.i.

Jeruzalémská 9, 116 52 Praha 1, Czech Republic

Authorized Body 235

Notified Body 1024



EC TYPE EXAMINATION CERTIFICATE No. 235/E-030/2015

This EC Type Examination Certificate is issued to:

Authorised representative: ONIKO–Europe, s.r.o., Pod Višňovkou 1662/21, 140 00 Praha 4 - Krč
Czech Republic

identification number: 24285081

Manufacturer: ONIKO Private Enterprise, ID 19242964
55/57, Voloska street, 04070 Kiev, Ukraine

PPE product: **One-side apron standard - model ON-RA 202**
One-side apron for children - model ON-RP 104
One-side surgical apron with shoulder straps - model ON-RA 203
One-side surgical apron with belt - model ON-RA 204
One-side surgical apron with velcro fastener - model ON-RA 205
One- side apron with shoulder straps – model ON-RA 214

It is certified that the manufacturer's technical file and above mentioned PPE product have been assessed and found to be in accordance with the basic health and safety requirements of Government Order 21/2003 Coll., which implements Council Directive No. 89/686/EEC "Personal Protective Equipment" as amended by Council Directives 93/68/EEC, 93/95/EEC and 96/58/EEC, as recorded in Report No.235/ZZ-028/2015, which is an integral part of this Certificate.

When examined the model was found to meet all of the relevant requirements of the appropriate harmonized standard(s):

ČSN EN 61331-3:1999 Protective devices against diagnostic medical X-radiation – Part 3: Protective clothing and protective devices for gonads
(Idt. IEC 61331-3:1998 a EN 61331-3:1999)

Marking and instructions have been assessed in the Czech language only.

Certificate is valid until 31.12.2020.

For and on behalf of Occupational Safety Research Institute
- Authorized Body 235, Notified Body No. 1024:

Date of Issue: 28.12.2015



Ing. Karel Škréta

Description and picture of the PPE product:



Apron is designated with shielding materials VINYL STANDARD or LEAD FREE or LEAD VINYL LIGHTWEIGHT or LEAD VINYL SUPERLIGHT for protection against X-Ray ionization in professional application according to instruction for using.



Graphical appearance of CE mark. In the event of the involvement of a notified body in the stage of manufacturing check its distinguishing number shall be added to CE mark.